

BIO-RAD Liquid Assayed Multiqual Level 1

For the control for clinical chemistry parameters of Element DC

FOR VETERINARY USE ONLY!

Antech

Smarter Diagnostics. Better Care.™



Item code 8185

Lot 46031

Expiry date

31.12.2027

Handling: Defrost and bring to room temperature.

Device	Parameter	conv.			SI		
		Target value	Target range	unit	Target value	Target range	unit
Element DC	GLU	65	44 – 86	mg/dl	3,6	2,4 – 4,8	mmol/L
Element DC	BUN	13,4	9,5 – 17,3	mg/dl	4,8	3,4 – 6,2	mmol/L
Element DC	CREA	0,7	0,4 – 1,0	mg/dl	61,9	35,4 – 88,4	μmol/L
Element DC	TP	4,2	3,0 – 5,4	g/dl	42	30 – 54	g/L
Element DC	ALB	2,3	1,7 – 2,9	g/dl	23	17 – 29	g/L
Element DC	AST	33	18 – 48	U/L	33	18 – 48	U/L
Element DC	ALP	55	16 – 94	U/L	55	16 – 94	U/L
Element DC	GGT	43	31 – 55	U/L	43	31 – 55	U/L
Element DC	TBIL	0,74	0,41 – 1,07	mg/dl	12,66	7,01 – 18,30	μmol/L
Element DC	CK	156	87 – 225	U/L	156	87 – 225	U/L
Element DC	Na	123	105 – 141	mmol/L	123	105 – 141	mmol/L
Element DC	K	2,6	2,0 – 3,2	mmol/L	2,6	2,0 – 3,2	mmol/L
Element DC	Cl	93	78 – 108	mmol/L	93	78 – 108	mmol/L
Element DC	PHOS	2,2	1,6 – 2,8	mg/dl	0,7	0,5 – 0,9	mmol/L
Element DC	Ca	6,6	4,5 – 8,7	mg/dl	1,6	1,1 – 2,2	mmol/L
Element DC	ALT	44	23 – 65	U/L	44	23 – 65	U/L
Element DC	CHOL	112	79 – 145	mg/dl	2,9	2,0 – 3,7	mmol/L
Element DC	TRIG	97	61 – 133	mg/dl	1,1	0,7 – 1,5	mmol/L
Element DC	AMY	57	36 – 78	U/L	57	36 – 78	U/L
Element DC	LIPA	40	25 – 55	U/L	40	25 – 55	U/L
Element DC	Mg	1,1	0,8 – 1,4	mg/dl	0,4	0,3 – 0,6	mmol/L